



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स) कल्याणी

All India Institute of Medical Sciences (AIIMS) Kalyani

(स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक सांविधिकनिकाय)

(A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)

राष्ट्रीय राजमार्ग - 34, बसन्तपुर, सागुना, कल्याणी, जिला - नदिया, पश्चिम बंगाल - 741245

NH-34 Connector, Basantapur, Saguna, Kalyani, District Nadia, West Bengal 741245

Ref: 367 /P-12017/1/22-23(E-Tender micro)

Date: 30.10.2024

E-Tender No. 309/P-12017/1/22-23(E-Tender micro)/ET-17

Dated: 25.09.24

1st Corrigendum of E-Tender for Chemiluminescence Based Fully Automated Immunoassay Analyzer for Infectious Disease Testing on Reagent Rental Basis at AIIMS, Kalyani

Following points are hereby modified as follows:

Sl. No.	Point	Written in original Tender Document	Correction after Pre-Bid meeting (to be read as)
1.	At page no. 7 of the E-tender document point 1	Analyzer should be capable to perform Anti HCV (with NS-5 Antigen)	Analyzer should be capable of performing Anti HCV with the latest technology as approved by WHO or US-FDA, CE-IVD and Indian CDSCO/SLA
2.	At page no. 7 of the E-tender document point 2	Should be Fully Automated random access Immunoassay System with a throughput of more than 160 tests/hour.	Should be Fully Automated random access Immunoassay System with a cumulative throughput of more than 150 tests/hour. Maximum 2 modules may be fixed by the Reagent Rental provider to fulfil the throughput requirement. Multiple smaller modules are not acceptable.
3.	At page no. 7 of the E-tender document point 6	Ability to change reagent, consumables & waste while instrument is performing the tests	DELETED
4.	At page no. 7 of the E-tender document point 8	Should have the capacity to load minimum 30 reagent pack in different parameters at a time.	Should have the capacity to load at least 15 reagent packs in different parameters at a time in each module – to a cumulative of 30 packs at a point of time.
5.	At page no. 7 of the E-tender document point 10	Should have facility for minimum 80 samples positions and separate STAT lane for emergency samples.	Should have facility for minimum 80 samples positions and separate STAT lane for emergency samples. Maximum 2 modules may be fixed by the Reagent Rental provider to fulfil the throughput requirement. Multiple smaller modules are not acceptable.
6.	At page no. 8 of the E-tender document point 16	More than 1500 sample tips can be loaded at a time to optimize the turnaround time & minimize human intervention during sample run. considering 160 tests/hour for 10 hours work shift.	More than 200 sample tips can be loaded at a time to optimize the turnaround time & minimize human intervention during sample run. considering 160 tests/hour
7.	At page no. 8 of the E-tender document point 18	Sample volume per test must ensure pediatric sample run with minimum volume requirement of 10 µl per test.	Sample volume per test must ensure pediatric sample run with minimum volume range of 10–100 µl per test.
8.	At page no. 23 of the E-tender document changes at Annexure VI (financial Bid) column (G)	Cost per test (Including Cleaner, Washer, Diluent, Kits, Controls, Calibrator, Qcs, Tips, required any other consumables etc.)	Cost per test (Including Cleaner, Washer, Diluent, Kits, Controls required as per NABL recommendation, Calibrator, Qcs, Tips, required any other consumables etc.)




Extension/01

E-Tender Ref. No: 309/P-12017/1/22-23(E-Tender micro) /ET-17, dt: 25.09.2024

E-Tender for Chemiluminescence Based Fully Automated Immunoassay Analyzer for Infectious Disease Testing on Reagent Rental Basis at AIIMS, Kalyani.

Sub: Extension of the Tender for another **15 days** from the date of this notification on the official Website of AIIMS, Kalyani and e-tender portal .

Consequent to approval of Competent Authority the above mentioned Tender for **Chemiluminescence Based Fully Automated Immunoassay Analyzer for Infectious Disease Testing on Reagent Rental Basis at AIIMS, Kalyani** is hereby extended for another **15 days** from the date of this notification on the official Website of AIIMS, Kalyani and e-tender portal.


Dr. Mahuya Chattopadhyay
Chairperson, Procurement

Copy to : i) PS to ED , ii) PS to DDA , iii) Indenting Department