

अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स) कल्यानी

All India Institute of Medical Sciences (AIIMS) Kalyani (स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक सांविधिकनिकाय) (A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI) राष्ट्रीय राजमार्ग – 34, बसन्तपुर, सागूना, कल्याणी, ज़िला – नदिया, पश्चिम बंगाल - 741245 NH-34 Connector, Basantapur, Saguna, Kalyani, District Nadia, West Bengal 741245

WEB CHALLENGE NOTICE

Ref.No: 409/BPS/PROC/NGEM/8/2024-BURNS&PS Date: 20/11/2024

The notice is being uploaded on the web site www.aiimskalyani.edu.in, e-wizard and CPPP.

Sub: Verification and Justification of proprietary nature of the items through 21 days WEB Challenge on the official website of AIIMS Kalyani, e-wizard and CPPP before Procurement of Tangential Hydro Surgical Debridement Device including two hand Pieces (Versa Jet II) for the Department of Burns & Plastic Surgery, AIIMS Kalyani against proprietary article certificate of the Original Equipment Manufacturer (O.E.M.) Inviting Comments thereon.

Department of Burns & Plastic Surgery, AIIMS Kalyani intends to buy Tangential Hydro Surgical Debridement Device (Versa Jet II) make by M/s Smith & Nephew Medical Limited (Address: 101, Hessle Road, HU3 2BN, England)

M/s Smith & Nephew Medical Limited (Address: 101, Hessle Road, HU3 2BN, England) declared in their company's letter head that M/s Smith & Nephew Healthcare Private Limited (Address: B-501-509, Dynasty Business Park, Andheri East, Mumbai-59, Maharashtra) is a 100% subsidiary of the OEM.

M/s Smith & Nephew Healthcare Private Limited, Mumbai, Maharashtra mentioned in their company's letter head that M/s Helios Medical Systems (Address: 105/1, Bidhan Nagar Road, Suncity Commercial Complex, Block-A, 2nd Floor, Kolkata -700067) is the authorized distributor of Tangential Hydro Surgical Debridement Device including two hand Pieces (Versa Jet II). Thus, these item(s) may be purchased from M/s Helios Medical Systems, Kolkata, India.

The Original Equipment Manufacturer (OEM) M/s Smith & Nephew Medical Limited (Address: 101, Hessle Road, HU3 2BN, England) has been declared in the Proprietary Certificate provided in their company's letter head that they are the sole manufacturer of the Tangential Hydro Surgical Debridement Device including two hand Pieces (Versa Jet II)

The notice is being uploaded for general information of prospective Manufacturer/ Authorized Distributor/Dealer to submit their objection/ comments, if any, on proprietorship of the Tangential Hydro Surgical Debridement Device including two hand Pieces (Versa Jet II), make: M/s Smith & Nephew Medical Limited (Address: 101, Hessle Road, HU3 2BN, England), they may submit their proposal along with specifications, supported by all documentary evidence and a price quotation for the above said items.

Continued from Page....01

The objection/proposal/comments, if any should be sent in sealed cover to the Office of the Chairperson, Procurement Cell, AIIMS, Kalyani, P.O.- Ghoragacha, NH-34 Connector, Basantapur, Saguna, Kalyani, District-Nadia, West Bengal- 741245. Or through email to web_challenge@aiimskalyani.edu.in_, so as to reach on or before 21 days from the web published date (before 5:00 pm) failing which it will be presumed that no other firm is interested to offer comments/protest/object and case will be decided on its merits.

The "Ref No. 409/BPS/PROC/NGEM/8/2024-BURNS&PS on or before 21 days from the web published date (before 5:00 p.m.) should be superscripted on sealed envelope.

Enclosures:

- 1) Proprietary Certificate of Manufacturer M/s Smith & Nephew Medical Limited (Address: 101, Hessle Road, HU3 2BN, England)
- 2) Authorization Certificate provided by the OEM to Indian Subsidiary M/s Smith & Nephew Healthcare Private Limited (Address: B-501-509, Dynasty Business Park, Andheri East, Mumbai-59, Maharashtra)
- 3) Authorization certificate provided by the Indian subsidiary to distributor: M/s Helios Medical Systems (Address: 105/1, Bidhan Nagar Road, Suncity Commercial Complex, Block-A, 2nd Floor, Kolkata -700067)
- 4) Specifications

Signed by Mahuya Chattopadhyay Date: 20-11-2024 17:14:38

Dr. Mahuya Chattopadhyay Chairperson Procurement Cell AIIMS, Kalyani

Copy to:

PS to ED
 For kind information please
 PS to DDA
 For kind information please
 For kind information please.
 Administrative Officer: For kind information please.
 FIC, Finance
 For kind information please.
 Indenting Officer

For kind information please
For kind information please

Smith & Nephew Medical Limited

101 Hessle Road Hull HU3 2BN England

T: + 44 (0)1482 225181 F: +44 (0)1482 328326 www.smith-nephew.com

27th February 2024

Proprietary Article Certificate

RE: VERSAJET II Hydrosurgery System

Smith+Nephew is the sole legal manufacturer and sole supplier of the Hydrosurgery System VERSAJET II globally.

The VERSAJET II Hydrosurgery System is a patented product of Smith+Nephew. The VERSAJET II Hydrosurgery System uses a high-pressure stream of sterile saline to optimise surgical debridement. The device enables surgeons to select, excise and evacuate non-viable tissue. Bacteria and contaminants from wounds, burns and soft tissue injuries using tissue preserving techniques.

The VERSAJET II Hydrosurgery System is unique and is protected by multiple international patents and patents pending. The VERSAJET II Hydrosurgery System is covered by or for use under one or more of the following U.S. Patents and their foreign counterparts:

Granted Patents:

- US8337175
- US8932269
- US8529498
- US9597107
- US9341184
- US10363061
- US11432838
- USD611961
- USD659165

We certify that the VERSAJET II System (console/Handpiece) is a proprietary product of Smith+Nephew.

DocuSigned by: Sam Atkinson Signer Name: Sam Atkinson Signing Reason: I approve this document Signing Time: 27-Feb-2024 | 14:19:00 GMT Signed by: 4247BB1FEDB542C29087D6B268F91213 Sam Atkinson

Name:


Principal Regulatory Affairs Specialist Position: Sums & Pla

Date: 27-Feb-2024 | 14:19:03 GMT aums Kalyani, WE CNB MCh MNAMS Assistant Professor. Eurns & Plastic Surgery AIIMS Kalyani, WB.

Smith Nephew

Registered No. 605496 in England and Wales, registered Office: PO Box 81, 101 Hessle Road, Hull, HU3 2BN, England Smith & Nephew Medical Limited, acting as agent for T.J. Smith and Nephew, Limited.

Wound Management Smith & Nephew Medical Ltd. 101 Hessle Road Hull HU3 2BN England T 44 (0) 1482 225 181 F 44 (0) 1482 328 326 www.smith-nephew.com



AUTHORISATION OF AGENT

I, the undersigned, hereby confirm that;

Smith & Nephew Healthcare Private limited, having its registered office at B-501-509, Dynasty Business Park, Andheri East, Mumbai – 59, Maharashtra, India are authorised agents for promoting and selling all Advanced Wound Management products of which Smith & Nephew Medical Limited, 101 Hessle Road, Hull, HU3 2BN, United Kingdom are the legal manufacturers.

Signed by:

Date:

10th April 2012

Name:

Liam Phelan

Position:

SVP Healthcare Systems

ry)



Smith & Nephew Medical Ltd

Registered No. 605496 in England and Wales at the above address

Smith & Nephew plc 15 Adam Street London WC2N 6LA England F +44 (0) 20 7401 7646 F +44 (0) 20 7930 3353 www.smith-nephew.com



TO WHOMSOEVER IT MAY CONCERN

We hereby declare that Smith & Nephew Healthcare Pvt. Ltd. having its registered office at B 501-509, Dynasty Business Park, Andheri Kurla Road, Andheri East, Mumbai – 400059 is a 100% subsidiary of Smith & Nephew Plc, its ultimate holding company.


Sincerely,

For Smith & Nephew Plc

Suar M. Swalez

Authorised Signatory





Political section is a second to the second second

What of Management

Certificate

This is to certify that Smith & Nephew Healthcare Pvt. Ltd. Is a wholly owned subsidiary of Smith & Nephew Group and is the sole authorised representative of Smith & Nephew Medical Limited products in India. Smith & Nephew India is the only authorised entity to quote for all Smith & Nephew Medical products in India and Smith & Nephew Medical Limited trades in India only through Smith & Nephew Pvt. Ltd. and its dealers.

Smith & Nephew Healthcare Pvt. Ltd has the authority to authorise any dealer to participate in any inquiry originating within India for Smith & Nephew Medical Limited products.

For Smith & Nephew Medical Limited

Signed by:

Name:

Liam Phelan

Position:

SVP Healthcare Systems

Date:

1st Dumber 2009



Smith & Nephew Healthcare Pvt Ltd

B-501-509, Dynasty Business Park Andheri Kurla Road, Andheri (E) Mumbai – 400059

CIN- U23239MH1994PTC080271

July 24, 2024.

TO.

THE EXECUTIVE DIRECTOR, AIIMS KALYANI, NH-34, BASANTPUR, NADIA WEST BENGAL 741 245

Subject: - <u>Authorization Letter of Versa jet (Hydro surgery debridement System)</u>

Dear Sir/ Madam

We hereby wish to inform you that for your esteemed hospital, the following Stockiest is authorized to deal on our behalf for our Advanced Wound Management (AWM) Division. The scope of the dealing would encompass quoting rates, negotiating, accepting and servicing orders for the AWM range of products of our company.

Tel: +91 22 4005 5090

Fax: +91 22 4005 5089

www.smith-nephew.com

M/S. HELIOS MEDICAL SYSTEMS (CONTACT PERSON: MR. SUMAN DEY) N05/I BINDHAN NAGAR ROAD. SUN CITY MARKET COMPLEX. BLOCK A. OFFICE NO F4. 2nd FLOOR. KOLKATA: - 700 067 PHONE NO.: +91- 09830170321 EMAIL: heliosmedicalsystems@gmail.com

This authorization is valid till formally rescinded from our side or till December 31, 2024 whichever is earlier.

In case of any further queries, please feel free to bring it to the knowledge of the undersigned.

Thanking you and Assuring you of our Best Services at all times.

Sincerely Yours

For Smith & Nephew Healthcare Private Limited

Komesh Teng Vice President – India Distributy

Dr. Sourabh Shankar Chakraborty
ONE MCh MNAMS
Assistant Professor,
Eurns & Plastic Surgery,
Alims Kalyani, WB.



Tender Specifications of Tangential Hydro Surgical Debridement Device

	Tangential Hydro Surgical Debridement Console
01.	It should have Hydrosurgery powered devise for surgical wound debridement producing high pressure (200-220 volt).
02.	It should have jet of sterile saline creating
03.	hand piece, and can provide sharp debridement and pulsed lavage irrigation It should have console (Front panel Layout: illuminated power switch Footswitch socket, Power Display).
04.	It should have Foot switch connection in the
05.	control, illuminated Green light Ring, Pump Interface, Key Lock Symbol). It should have Foot switch pedal (Weight: 3lbs/1.1 Kg (maximum), Cord Length 15 ft/4.6 mtr [minimum].
06.	It should be CE and USFDA Certified.
07.	Warranty should be at least for 05 (five) Years.
	Hand Pieces
01.	It should have angled 01 (one) handpiece per console 45 degree with a cutting window of 8mm with high pressure cord and disposed to the control of the contr
)2.	window of 8mm with high pressure cord and disposable tubing. It should have angled 01 (one) handpiece per console 45 degree with a cutting window of 14mm with high pressure cord and disposable tubing.



