



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स) कल्याणी

All India Institute of Medical Sciences (AIIMS) Kalyani

(स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक सांविधिकनिकाय)

(A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)

राष्ट्रीय राजमार्ग – 34, बसन्तपुर, सागूना, कल्याणी, ज़िला – नदिया, पश्चिम बंगाल - 741245

NH-34 Connector, Basantapur, Saguna, Kalyani, District Nadia, West Bengal 741245

Ref No :427/D-39017/1/23(HITES)01

Dated :26/11/2024

E-Tender No – 160/D-39017/1/23/HITES/01/ET-10

3rd corrigendum of E-tender for procurement of 4K Arthroscope with camera system, shaver system, fluid management system and complete knee instrumentation set(ACL & PCL set)

SL.NO	Existing Para No. /Page No/ Clause No.	Contents existing	Contents Revised/To be read as	Remarks
1	Page no. 02 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	A full 4K high definition processor should have native resolution of 3840 x 2160 pixels system .	A full 4K high definition processor should have native resolution of 3840 x 2160 pixels system or A full 3 Chip 4K UHD processor should have native resolution of 3840 x 2160 pixels system.	For wider participation
2	Page no. 02 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Should also provide 3840 x 2160 or 1920 x 1080 should also be selectable.(Should provide 4K output).	Should provide resolution of 3840 x 2160.	For wider participation
3	Page no. 03 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Powerful LED Light source- Qty-01	Same as existing	Same as existing
4	Page no. 03 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Should have various input/output terminals, including 3G/HD/SD SDI, DVI-D, BNC (x5) and DP.	Should have various input/output terminals, including 3G/HD/SD SDI, DVI-D, BNC (x5), DP & HDMI.	For wider participation

5	Page no. 04 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Should be a 4K UHD/full HD.	Same as existing .	Same as existing
6	Page no. 04 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Recording media: Internal HDD(100 GB OR More) and external USB port for storage.	Recording media: Internal HDD(100 GB OR More) and external USB port for storage which is optional.	For wider participation
7	Page no. 05 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	The system should be BIS or European CE or US FDA certified.	The system should be BIS or European CE or US FDA certified or CDSCO.	For wider participation
8	Page no. 07 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	The Unit should have 2 modes for Normal and aggressive Reaction so as to balance efficacy with safety.	The Unit should have 2 modes for Normal and aggressive Reaction so as to balance efficacy with safety Which is optional.	For wider participation
9	Page no. 09 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Basket Punch slim, Bite width – 1.5mm – 2mm.	Basket Punch slim, Bite width – 1.5mm – 3mm.	For wider participation
10	Page no. 09 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Posterior/ Reverse Punch Upbiter, Bite width- 2 – 3mm.	Posterior/ Reverse Punch Upbiter, Bite width- 2 – 3mm which is optional.	For wider participation
11	Page no. 09 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Posterior/ Reverse Punch Straight, Bite width- 2 – 3mm.	Posterior/ Reverse Punch Straight, Bite width- 2 – 3mm which is optional.	For wider participation
12	Page no. 09 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Thickness tester	Thickness tester or ACL graft sizing block.	For wider participation

13	Page no. 10 of 1st corrigendum ref.no.:225/D-39017/1/23 (HITES) 01 Dt. 07.08.2024	Currecte	Currecte (which is optional).	For wider participation
14	Page No 4,5,8,23 of the tender document	'Performance Security' will be 10% of the value of contract.	'Performance Security' will be 5% of the value of contract.	

Consequent to the approval of the competent authority, the above-mentioned tender is hereby extended for 14 days from the date of this notification on the official website of AIIMS Kalyani and CPPP.

(Prof.) Dr. Mahuya Chattopadhyay
Chairperson of Procurement Cell
AIIMS Kalyani