



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स) कल्याणी

All India Institute of Medical Sciences (AIIMS) Kalyani

(स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक सांविधिकनिकाय)

(A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)

राष्ट्रीय राजमार्ग - 34, बसन्तपुर, सागूना, कल्याणी, ज़िला - नदिया, पश्चिम बंगाल - 741245

NH-34 Connector, Basantapur, Saguna, Kalyani, District Nadia, West Bengal 741245

WEB CHALLENGE NOTICE

Ref.No: ⁶³⁹ /OPTLGY/4/2024-OPHMLGY/WC- 08

Date: 14.02.2025

The notice is being uploaded on the web site www.aiimskalyani.edu.in, e-wizard and CPPP.

Sub: Verification and Justification of proprietary nature of the items through 21 days Web Challenge on the official website of AIIMS Kalyani, e-wizard and CPPP before Procurement of "Consumables (B Scan Probe) for Eyecubed, Model: B Scan System, Ellex" for the Department of Ophthalmology, AIIMS Kalyani against proprietary article certificate of the Original Equipment Manufacturer (O.E.M.) Inviting Comments thereon.

Department of Ophthalmology, AIIMS, Kalyani intends to buy "Consumables (B Scan Probe) for Eyecubed, Model: B Scan System", make by M/s Ellex Medical Pty Ltd. (Address: 2-4 Second Avenue Mawson Lakes, SA 5095, Australia).

M/s Ellex Medical Pty Ltd. (Address: 2-4 Second Avenue Mawson Lakes, SA 5095, Australia) mentioned in the authorization letter provided in their company's letter head that M/s Toshbro Medicals Pvt. Ltd. (Address: Khetan Bhavan, 198 Jamshedji Tata Road, Churchgate, Mumbai 400020, India) is the authorized distributor. Thus, these consumables may be purchased from M/s Toshbro Medicals Pvt. Ltd. Mumbai.

M/s Ellex Medical Pty Ltd. Australia has been declared in the Proprietary Article Certificate provided in their company's letter head that they are the sole manufacturer of the "Consumables (B Scan Probe) for Eyecubed, Model: B Scan System, Ellex".

The notice is being uploaded for general information of prospective Manufacturer/ Authorized Distributor/Dealer to submit their objection/ comments, if any, on proprietorship of the "Consumables (B Scan Probe) for Eyecubed, Model: B Scan System", Make: M/s Ellex Medical Pty Ltd. (Address: 2-4 Second Avenue Mawson Lakes, SA 5095, Australia), they may submit their proposal along with specifications, supported by all documentary evidence and a price quotation for the above said items.

Meh



Continued from Page....01

The objection/proposal/comments, if any should be sent in sealed cover to the Office of the Chairperson, Procurement Cell, AIIMS, Kalyani, P.O.- Ghoragacha, NH-34 Connector, Basantapur, Saguna, Kalyani, District-Nadia, West Bengal- 741245. Or through email to web_challenge@aiimskalyani.edu.in, so as to reach on or before 21 days from the web published date (before 5:00 pm) failing which it will be presumed that no other firm is interested to offer comments/protest/object and case will be decided on its merits.

The "Ref No. **639 /OPTLGY/4/2024-OPHMLGY/WC-08** on or before 21 days from the web published date (before 5:00 p.m.) should be superscripted on sealed envelope.

Enclosures:

- 1) Proprietary Certificate of Manufacturer **M/s Ellex Medical Pty Ltd. (Address: 2-4 Second Avenue Mawson Lakes, SA 5095, Australia)**
- 2) Authorization Certificate provided by the OEM to **M/s Toshbro Medicals Pvt. Ltd. (Address: Khetan Bhavan, 198 Jamshedji Tata Road, Churchgate, Mumbai 400020, India)**
- 3) Technical Specification

Moh
14/2/25
Chairperson/Co-chairperson of Procurement Cell
(AIIMS Kalyani)

Copy to:

1. PS to ED : For kind information please
2. PS to DDA : For kind information please
3. FIC Website : For kind information please.
4. Administrative Officer : For kind information please.
5. FIC, Finance : For kind information please.
6. Indenting Officer : For kind information please





Ellex Medical Pty Ltd
3-4 Second Avenue
Mawson Lakes, SA 5095
AUSTRALIA
ABN 35 008 276 060
T: +61 8 7074 8200

TO WHOM IT MAY CONCERN

Date: 24 September 2024

Proprietary Certificate

This is to certify that we Ellex Medical Pty Ltd, 3-4 Second Avenue, Mawson Lakes, SA 5095, Australia are the manufacturer of the Eyecubed Ultrasound AB Scan and their Probes and Spares.

The Probes and Spares manufactured by Ellex Medical Pty Ltd are proprietary items of our Company Ellex Medical Pty Ltd, Australia and no other company can supply the same.

Best regards,

For Ellex Medical Pty Ltd


Digitally signed by
Angela Standing
Date: 2024.09.24
13:23:55 +09'30'

Angela Standing
General Manager

Ellex Medical Pty Ltd
Lumibird Medical Group



Ellex Medical Pty Ltd
3-4 Second Avenue
Mawson Lakes, SA 5095
AUSTRALIA
ABN 35 008 276 060
T: +61 8 7074 8200

Date: 20 December 2024

MANUFACTURER'S AUTHORIZATION LETTER

TO WHOM IT MAY CONCERN

We, Ellex Medical Pty Ltd, 3-4 Second Avenue, Mawson Lakes, SA 5095, Australia who are established and reputable manufacturer of Ophthalmic Medical Equipment do hereby authorize


Toshbro Medicals Pvt. Ltd.
Khetan Bhavan,
198 Jamshedji Tata Road,
Churchgate, Mumbai 400 020
India
Tel. 022-4355 2727

to submit a bid, and subsequently negotiate and sign the Contract with you for the goods manufactured by us.

No Company or Firm or Individual other than M/s. Toshbro Medicals Pvt. Ltd., are authorized to bid, negotiate and conclude the contract in regard to this business.

This letter is valid until 31st December 2025 unless revoked beforehand.

SIGNED BY AN AUTHORISED REPRESENTATIVE OF ELLEX MEDICAL PTY LTD



Digitally signed by
Angela Standing
Date: 2024.12.20
09:16:02 +10'30'

.....
Angela Standing
General Manager

Ellex Medical Pty Ltd
Lumibird Medical Group

B Scan Probe- Technical Specifications

1. Should be of 10 MHz frequency
2. Should be for the posterior segment
3. Should be able to differentiate membranes such as posterior vitreous detachment (PVD) in opaque Media.
4. Should be able to differentiate intraocular and orbital tumours
5. Should be a sealed probe
6. Should have a focal range of 15-35 mm
7. Should have a focal depth of 25 mm
8. Image width at focal zone: 19-36 mm
9. Scanning angle: 52 degrees
10. Probe should be compatible with Ellex Eyecubed AB Scan system


24/10/25

Dr. ABHISHEK
Associates Professor
Vitreous Retina Surgeon
Department of Ophthalmology
AIIMS Kalyani, Pin-741245