



**अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स) कल्याणी**  
All India Institute of Medical Sciences (AIIMS) Kalyani  
(स्वास्थ्य एवं परिवार कल्याण मंत्रालय, भारत सरकार के तत्वावधान में एक सांविधिकनिकाय)  
(A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)  
राष्ट्रीय राजमार्ग – 34, बसन्तपुर, सागूना, कल्याणी, ज़िला – नदिया, पश्चिम बंगाल - 741245  
NH-34 Connector, Basantapur, Saguna, Kalyani, District Nadia, West Bengal 741245

---

Ref No: 331/RCC/PROC/NGEM/2/2024-RCC (COMP – 1659)

Dated: 19.08.2025

E-Tender No: 54/RCC/PROC/NGEM/2/2024-RCC (COMP – 1659)/ET-04

Dated: 29.04.2025

**Corrigendum Notice**

**Corrigendum Notice of E-Tender of Rate Contract for Supply of Consumables/Non-Consumables  
for Cardiac Cath Lab in AIIMS Kalyani for Department Of Cardiology**



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), कल्याणी  
**All India Institute of Medical Sciences (AIIMS), Kalyani**  
 (स्वास्थ्यएवंपरिवार कल्याण मंत्रालय, भारत सरकार के तत्वाबधान में एक संबिधिकनिकाय)  
 (A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)  
 राष्ट्रीय राजमार्ग-34, बसंतपुर, सगुना, कल्याणी, नदीया, पश्चिम बंगाल-741245  
 NH-34 connector, Basantpur, Saguna, Kalyani, Nadia, West Bengal -741245  
 Website: [www.aiimskalyani.edu.in](http://www.aiimskalyani.edu.in)  
 E-Tendering Portal: <https://www.tenderwizard.com/AIIMSKALYANI>

Corrigendum-01

**E-TENDER DOCUMENT** : Tender for supply of Consumables / Non Consumables for Cardiac Cath Lab  
**E-Tender No.** : 54/RCC/PROC/NGEM/2/2024-RCC (COMP – 1659)/ET-04

<b>Sl. No.</b>	<b>Reference</b>	<b>Written As under the heading of "General Terms and Conditions"</b>	<b>To be Read As (under the heading of "General Terms and Conditions")</b>	<b>Remarks of the RCC justifying the modifications</b>
1	Section –III Eligibility Criteria	3.1 Bidder shall only be a manufacturer having valid own manufacturing license with product registration certificate issued by the Drugs Controller General of India wherever applicable	3.1 Bidder <b>can</b> be a manufacturer having valid own manufacturing license with product registration certificate issued by the Drugs Controller General of India wherever applicable	To allow maximum participation.
2	Mode of Procurement		<ul style="list-style-type: none"> <li>• Items Costing Below ₹ 5000 will be purchased by AIIMS, Kalyani on Quarterly basis as per the requisition of the end user.</li> <li>• Items costing above ₹ 5000 will be purchased by AIIMS, Kalyani as and when required on consignment basis.</li> </ul>	To keep procurement process simple. The high cost consumables are usually provided on consignment by vendors.
3	For Non Drug Eluting Stents-		One product of lowest quote of each certification viz: <ul style="list-style-type: none"> <li>• USFDA Approved</li> <li>• CE Approved</li> <li>• Manufactured in India/DGCI Approved</li> </ul> Will be taken as and when required on Consignment Basis.	This will provide more option to patients and faculty for use of devices like stent. This will also help serve patient's from varying economic strata.
4	For Drug Eluting Stents		One product of lowest quote of each certification viz: <ul style="list-style-type: none"> <li>• USFDA Approved</li> <li>• CE Approved</li> </ul>	This will provide more option to patients and faculty for use of devices like stent. This will



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), कल्याणी  
**All India Institute of Medical Sciences (AIIMS), Kalyani**  
 (स्वास्थ्यएवंपरिवार कल्याण मंत्रालय, भारत सरकार के तत्वाबधान में एक संबिधिकनिकाय)  
 (A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)  
 राष्ट्रीय राजमार्ग-34, बसंतपुर, सगुना, कल्याणी, नदीया, पश्चिम बंगाल-741245  
 NH-34 connector, Basantpur, Saguna, Kalyani, Nadia, West Bengal -741245  
 Website: [www.aiimskalyani.edu.in](http://www.aiimskalyani.edu.in)  
 E-Tendering Portal: <https://www.tenderwizard.com/AIIMSKALYANI>

Sl. No.	Reference	Written As under the heading of "General Terms and Conditions"	To be Read As (under the heading of "General Terms and Conditions")	Remarks of the RCC justifying the modifications
			<ul style="list-style-type: none"> <li>Manufactured in India/DGCI Approved</li> </ul> <p>Will be taken as and when required on Consignment Basis.</p>	also help serve patient's from varying economic strata.
5	For Pacemaker		<p>One product of lowest quote of each certification viz:</p> <ul style="list-style-type: none"> <li>USFDA Approved</li> <li>CE Approved</li> <li>Manufactured in India/DGCI Approved</li> </ul> <p>Will be taken as and when required on Consignment Basis.</p>	This will provide more option to patients and faculty for use of devices like pacemaker.
6	Section-IX List of Consumable Items along with Specification	<b>17. INFLATION DEVICE :</b> BIS/ USFDA & CE or Japan PMDA Approved/Certified, Inflation device with manometer upto <b>30 atmosphere</b> (easy to operate with luminescent dial) Gun type	<b>17. INFLATION DEVICE :</b> <b>INFLATION DEVICE : BIS/ USFDA &amp; CE or Japan PMDA Approved/Certified, Inflation device with manometer upto 26 atmosphere</b> (easy to operate with luminescent dial) Gun type	This will increase participation.
7	Section-IX List of Consumable Items along with Specification	<b>For S. No. 201r / 201s/ 201t / 201u</b> ATP therapies to terminate high rate atrial tachyarrhythmia episodes.	<b>For S. No. 201r / 201s/ 201t / 201u</b> AT/AF Management.	This will keep specification more generic.
		<b>For S. No. 202B/ 202C/ 202D/ 202G/ 202H/ 202I/ 202J/ 202K/ 202L/ 202M/ 202N/ 202O</b>	<b>For S. No. 202B/ 202C/ 202D/ 202G/ 202H/ 202I/ 202J/ 202K/ 202L/ 202M/ 202N/ 202O</b> Should <b>PatientNotifier(Audible/Vibratory)</b>	This will keep specification more generic.



अखिल भारतीय आयुर्विज्ञान संस्थान (एम्स), कल्याणी  
All India Institute of Medical Sciences (AIIMS), Kalyani  
(स्वास्थ्यएवंपरिवार कल्याण मंत्रालय, भारत सरकार के तत्वाबधान में एक संबिधिकनिकाय)  
(A Statutory Body under the Aegis of Ministry of Health and Family Welfare, GOI)  
राष्ट्रीय राजमार्ग-34, बसंतपुर, सगुना, कल्याणी, नदीया, पश्चिम बंगाल-741245  
NH-34 connector, Basantpur, Saguna, Kalyani, Nadia, West Bengal -741245  
Website: [www.aiimskalyani.edu.in](http://www.aiimskalyani.edu.in)  
E-Tendering Portal: <https://www.tenderwizard.com/AIIMSKALYANI>

Sl. No.	Reference	Written As under the heading of "General Terms and Conditions"	To be Read As (under the heading of "General Terms and Conditions"	Remarks of the RCC justifying the modifications
		<p>Should <b>audibly alert</b> the patient in case of RV Noise &amp; Lead failure</p> <p>Should monitor the fluid build-up status in a heart failure patient and <b>audible alert</b> the patient in case its critical</p> <p><b>For 202I/ 202J/ 202K/ 202L</b></p> <p>Should have facility for active fixation of LV lead in CS with both screwing and/or deployable lobes with anchor mechanism</p>	<p>the patient in case of RV Noise &amp; Lead failure</p> <p>Should monitor the fluid build-up status in a heart failure patient and <b>PatientNotifier (Audible/Vibratory)</b> the patient in case its critical</p> <p><b>For 202I/ 202J/ 202K/ 202L</b></p> <p>Should have facility for active or passive fixation of LV lead in CS.</p>	This will keep specification more generic.